

CH. PERVAIZ ELAHI INSTITUTE OF CARDIOLOGY, MULTAN

TECHNICAL COMPARATIVE STATEMENT FOR THE YEAR 2015.16

NAME OF EQUIPMENT: OPERATING LIGHT WITH LED MOBILE TYPE (Part-B)

Model	M/S Allimed Solution	M/S Eastern Medical Technologies	Sahar International
1			
2	marLED E9j	Qe 60	LED 300 DF with spect
3	Gebruder Martin Germany	Brandon UK	Dr. Mach Germany
4	31.12.2015	31.12.2019	01.12.2017
5	CE	CE	CE
6	5 Years	5 Years	5 Years
7	Mobile Type	Mobile Type	Mobile Type
8	40,000 hour	60,000 hour	50,000 hour
9	160,000 lux	160,000 lux	130,000 lux
10	CRI 95	CRI 95	CRI 95
11	3600 to 4600 Kelvin	3100 to 5000 Kelvin	4500 Kelvin
12	16 to 23cm	20 to 42cm	14 to 28cm
13	Sterilizable handle	Sterilizable handle	Sterilizable handle
14	Maintenance free	Maintenance free	Maintenance free
15	Automatic	Automatic	Automatic
16	2 Hour	2 Hour	3 Hour
17	220/50Hz	220/50Hz	220/50Hz
REMARKS	TA	TA	TA

Dr. Haider Zaman  
Prof. of Cardiac Surgery  
CPEIC, Multan  
(Chairman)

Dr. Tariq Abbas  
Associate. Prof. of Cardiology  
CPEIC, Multan  
(Member)

Dr. Rana Altaf  
Prof. of Anesthesia  
CPEIC, Multan  
(End-User)

Dr. Sohail Saleemi  
Sr. Registrar Cardiology  
CPEIC, Multan  
(Member)

Dr. Muhammad Moeen  
Sr. Cardiac Surgery  
CPEIC, Multan  
(Member)

Dr. Fayyaz Hashmi  
Sr. Medical Officer  
CPEIC, Multan  
(Member)

Dr. Sohail Ahmad  
Assist. Prof. Anaesthesia  
CPEIC, Multan  
(Member)

Dr. Farhad Khan  
Director Technical  
CPEIC, Multan  
(Member)

Engr. Muhammad Kaleem  
Biomedical Engineer  
CPEIC, Multan  
(Member)